

## 資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和7年2月14日

防衛省防衛医科大学校病院

事務部長 小笠原 誠

◎調達機関番号 010 ◎所在地番号 11

### 1 調達内容

(1) 品目分類番号 22, 31

(2) 導入計画物品及び数量

アーム型X線CT診断装置 一式

(3) 調達方法 購入等

(4) 導入予定時期

令和8年度第3・四半期以降

(5) 調達に必要とされる基本的な要求要件

① システムの構成

アーム型X線CT診断装置 1 式、

全標準付属品及び全オプション付属品を備えること。

② システムの性能

(A) 装置移動が可能で手術中に使用できること。

- (B) ボクセルサイズ80  $\mu$ m以下の超高解像度高画質であること。
- (C) 画像の3D再構成が可能であること。
- (D) トモシンセシス 撮影が可能であること
- (E) 頭部、手、足を撮影できること。
- (F) 既存の画像保管通信システム(PACS)、病院情報システム(HIS)、放射線情報システム(RIS)に接続できること。

③ その他

- (A) 操作性及び安全性に優れていること。
- (B) 将来の機能拡張に対応できること。

2 資料及びコメントの提供方法 上記1(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

- (1) 資料等の提出期限 令和7年4月4日17時00分  
(郵送の場合は必着のこと。)
- (2) 提出先 〒359-8513 埼玉県所沢市並木3-2  
防衛省防衛医科大学校病院事務部病院運営課

運営管理係 村田 圭

電話04-2995-1511（内線3017）

3 その他 本公表内容は予定であり、変更することがあり得る。

4 Summary

(1) Classification of the products to be procured : 22, 31

(2) Nature and quantity of the products to be purchased : Arm-Type X-ray CT Diagnostic System 1Set

(3) Type of the Procurement : purchase

(4) Time limit for the submission of the requested material : 17:00 4, April, 2025

(5) Basic requirements of the Procurement

① System configuration:

It should have aX-ray CT diagnostic system, all standard attachments and all optional attachments.

② System performance:

(A) The device should be mobile and operable

during surgery.

(B) It should provide ultra-high resolution and high-quality imaging with a voxel size of 80  $\mu\text{m}$  or less.

(C) The system should support 3D image reconstruction.

(D) Tomosynthesis imaging should be supported.

(E) It should be capable of imaging the head, hands, and feet.

(F) The system should be compatible with existing Picture Archiving and Communication Systems (PACS), Hospital Information Systems (HIS), and Radiology Information Systems (RIS).

③ Others

(A) It should be highly safe and easy to be operated.

(B) It should be able to be upgraded in the future.

(6) Contact point for the notice : Kei Murata,  
Hospital Management Division, Administrative

Department, National Defense Medical College  
Hospital, 3-2 Namiki Tokoro-  
zawashi Saitama 359-8513 Japan  
TEL 04-2995-1511 (ext. 3017)