

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和7年2月14日

防衛省防衛医科大学校病院

事務部長 小笠原 誠

◎調達機関番号 010 ◎所在地番号 11

1 調達内容

(1) 品目分類番号 22, 31

(2) 導入計画物品及び数量

人工心肺装置 一式

(3) 調達方法 購入等

(4) 導入予定時期

令和8年度第3・四半期以降

(5) 調達に必要とされる基本的な要求要件

① システムの構成

人工心肺装置1式、

全標準付属品及びバックアップ機器を備えること。

メインポンプの装置トラブルに速やかに対応できるシ

ステム構成であること。

急な大量出血や回路内残血処理に対応できるシステム構成であること。

体外循環中の血液ガスを簡便かつ連続的にモニタリングできるシステム構成であること。

② システムの性能

(A) 送脱血流量連動システムを搭載していること。

(B) 既存の麻酔記録装置（ERGA）と接続できること。

(C) 酸素供給量（D02）管理のサポート機能があること。

③ その他

(A) 操作性及び安全性に優れていること。

(B) 将来の機能拡張に対応できること。

2 資料及びコメントの提供方法 上記 1 (2) の物品に関する一般的な参考資料及び同 (5) の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1) 資料等の提出期限 令和7年4月4日17時00分
(郵送の場合は必着のこと。)

(2) 提出先 〒359-8513 埼玉県所沢市並木 3-2
防衛省防衛医科大学校病院事務部病院運営課
運営管理係 村田 圭
電話04-2995-1511 (内線3017)

3 その他 本公表内容は予定であり、変更することがあり得る。

4 Summary

- (1) Classification of the products to be procured : 22, 31
- (2) Nature and quantity of the products to be purchased : Heart-lung Machine System 1Set
- (3) Type of the Procurement : purchase
- (4) Time limit for the submission of the requested material : 17:00 4, April, 2025
- (5) Basic requirements of the Procurement
 - ① System configuration:

It should have a Heart-lung Machine System, all standard attachments and backup equipment.

The system configuration must enable rapid

response to any main pump equipment trouble.

The system configuration must be able to handle sudden massive bleeding and blood residuals in the circuit.

The system configuration must allow for easy and continuous monitoring of blood gases during extracorporeal circulation.

② System performance:

(A) The system should be equipped with an interlocking system for sending/receiving blood flow.

(B) Ability to connect to existing anesthesia recording equipment (ERGA).

(C) The system must have the ability to support oxygen supply (D02) management.

③ Others

(A) The system shall be easy to operate and safe.

(B) The system should be able to accommodate future enhancements.

(6) Contact point for the notice : Kei Murata,

Hospital Management Division, Administrative
Department, National Defense Medical College
Hospital, 3-2 Namiki Tokoro-
zawashi Saitama 359-8513 Japan
TEL 04-2995-1511 (ext. 3017)